

Traceable® Products are ISO 9001:2018 Quality-
Certified by DNV and ISO/IEC 17025:2017
accredited as a Calibration Laboratory by A2LA.

©2020 Traceable® Products. 92-1062-10 Rev. 5 041720

**60-MINUTE
INTERVAL TIMER
INSTRUCTIONS**

SPECIFICATIONS

Range: 0 to 60 minutes
Resolution: 1 minute
Alarm: Audible
Size: $3\frac{3}{4}$ H x $3\frac{3}{4}$ W x 2D
inches
Weight: $3\frac{3}{4}$ ounces

OPERATION

Turn the knob clockwise to 60
and then turn it back to the
desired time.

The bell will ring at the end of the
selected time.