













# Insta-Test<sup>®</sup> ANALYTIC

## Free & Total Chlorine

25 Test Strips


Free Chlorine ppm						
FCI						
	0	0.5	1	3	5	10

Total Chlorine ppm						
TCI						
	0	0.5	1	3	5	10

Immerse strip for **2 seconds**.  
Remove strip with **pads face up**.  
**SHAKE ONCE TO REMOVE EXCESS WATER.**

Read immediately: FCI > TCI.



- Keep wet fingers out of the bottle.
- Cap bottle tightly after removing strip.
- Store in a cool, dry place.