

# **TABmaster Model 6715**

# **Instruction Manual**



Be sure to read this manual thoroughly before using the instrument, fully understand and pay attention to each cautions mentioned. Please well-keeping this manual for long time service reference.



Product Configuration
Caution5
1. Introduction
1.1. Features
1.2. Specifications
2. Outlook & Structure 8
2.1. Base structure 8
2.2. Micromanometer 9
3. Settubg up9
3.1. Capture Hood9
<b>3.2. Velocity Grid</b>
3.3. Pitot tube
3.4. Micromanometer
3.5. Shoulder Strap
4. Operating Instruction
<b>4.1. Power supply</b>
4.2. Start On / Power Off
4.3. Keypad operation
5. Function test
5.1. User Interface
<b>5.2. Single Mode</b>
5.3. Average Mode
5.4. Back Pressure Mode
6. Menu setting
6.1. General setting
<b>6.1.1. Date</b>
<b>6.1.2. Time</b>
<b>6.1.3. Auto Off</b>
6.1.4. Back lighting
6.1.5. Communication
<b>6.2. Test setting</b>

6.2.2. ID application	20
6.2.3. Test Mode	20
6.2.5. Data record	21
6.2.6. Standard/Actual test	21
6.2.7. Compensation factor	22
7 Error and Troubleshooting	23
8 Warranty and Service	24
8.1. Product Warranty	24

# 6715 Standard:

Item.	Qt'y
Micromanometer Model 6700	1
Instrument Base	1
Fabric Hood: 2ft×2ft	1
Carrying case	1
Portable handle	1
Frames	1
Poles	4
Communication cable	1
Calibration certificate	1
CD-Rom(Manual and software inside)	1
Quick-Guide	1

# 6715 Optional Parts:

Item.	Specifications
Spare hood 2x2ft	
Spare hood 2x4ft	
Spare hood 1x4ft	
Spare hood 3x2ft	
Spare hood 3x3ft	
Spare hood 500x500mm	
Poles	
Capture Hood Stand	Up to 3.4m with capture hood
Communication cable	
Printer	
AC adapter	DC5V
International plug adapter	
Carrying case	
Velocity Grid	
Pitot tube (2.3 x 200mm)	
Pitot tube (4 x 300mm)	
Pitot tube (8 x 500mm)	
Pressure tubing	2*250cm

Item.	Qt'y
Micromanometer Model 6700	1
Carrying case	1
Pressure tubing	2
Communication cable	1
Calibration certificate	1
CD-Rom(Manual and software inside)	1
Quick-Guide	1

# 6700 Optional Parts:

Item.	Specifications
Velocity Grid	
Pitot tube (2.3 x 200mm)	
Pitot tube (4 x 300mm)	
Pitot tube (8 x 500mm)	
Pressure tubing	2*250cm
Shoulder Strap	
Printer	
AC adapter	DC5V
International plug adapter	
Communication cable	

# Caution.

The symbols for warning mentioned in this manual are defined below:

# (symbols classifications)



Danger: To Prevent Serious Injury or Death



 $Warnings\ in\ this\ dassification\ indicate\ a\ danger\ that\ may\ result in\ serious\ injury\ or\ death\ if\ not\ observed\ .$ 

# **Caution:** To Prevent Damage to the Product

Warnings in this dassification indicate a risk of damage to the product that may void the product warranty if not observed.

# ( Description of Symbols)



 $\Delta$  This symbol indicates a condition that requires caution (induding danger). The subject of each caution is illustrated inside the triangle.



This symbol indicates a prohibition. Do not take the prohibited action shown inside or near this symbol.



This symbol indicates a mandatory action. A specific action is given near the symbol.

<b>⚠ WARNING</b>		
	0	Never bring the fabric hood to a flammable gas
		a tmosphere.
Heated forbidden	•••	otherwise, the heat may cause a fire or explosion.
	0	Do not disassembly or refit the instrument.
Disassemble		otherwise, may cause the electric shock or a fire.
prohi bitio n		
0	0	Using properly under the instruction manual.
Using properly	•••	otherwise, may cause sensor damaged or an electric shock even a fire.
	0	If abnormal smells, noises or smoke occur, or if liquid enters the instrument, pull out
		the AC adapter and remove the batteries immediately. Then send it to the
		maintenance Dept. of KANOMAX for after service.
Using properly	•••	or, there is possible of an electric shock or a fire or instrument malfunction.
	0	Do not expose the fabric hood, base and the instrument
(\)		noumenon in the rain.
	•••	otherwise, may cause an electric shock, an fire and an
Forbidde n		person injure.

			ZIX CAUTION
		0	Always unplug when the instrument not in use.
	Using properly		Failure to do so may cause an electric shock, an fire or circuit da mage.
	0	0	Remove the batteries when storing the instrument for a long period. Do not leave the exhaused batteries in the battery comparement and exchange the batteries on time.
	Using properly		Failure to do so may cause the batteries leakage and the instrument damage.
	0	0	Matching the correct AC adapter outward as reqested.
	Using properly		or may cause the instrument damage.
		0	Do not wipe the instrument with a volatile solvent.
	$\bigcirc$	•••	the body may deform or deteriorate. Use a soft dry doth to remove stains. If stains persist, soak the doth in a neutral detergent and wipe the instrument with the soft
L	Forbidde n		cloth. Never use volatile solvents such as thinner or benzene.
	0	0	Do not use or leave the instrument in a high temperature, high humidity or dusty environment. Do notleave the instrument under direct sunlight.
		•••	otherwise, the instrument may not function properly
	Forbidde n		out of the specified operating conditions or the inside components damaged.
	$\bigcirc$	0	Never dropping the unit or placing heavy objects on it

It may cause damage or malfunction to the instrument

Forbidde n

L. IIICI Odaccioii

TABmaster Model 6715 is a kind of intelligent test instrument with multi functional of airflow test, velocity test and micro-Differential pressure test. It's widely used in air flow and velocity testing of air - conditioning, HAVC system and other places, especially for the high precision micro differential pressure test.

#### 1.1. Features

- 3.5 inch true color LCD.
- Available measuring the air flow, velocity, temperature and humidity at the same time.
- Two kinds of tests are available of in flowing and exhaust flowing.
- Back pressure compensation.
- Ultra-large Storage capacity: 8000 records
- ♦ Blue tooth communication makes remote monitoring and data transferring available.
- ◆ Light weight, compact configuration for easy carrying.
- Data printing or transferring to the computer.


# 1.2. Specifications

Items		Specifications
Air flow	Test range	$40{\sim}4300{\rm m}^3{ m /h}$
	Accuracy	$\pm 3\%$ of readings $\pm 8$ m <sup>3</sup> /h (>85m <sup>3</sup> /h)
	Resolution	1 m³ /h
	Test range	$0.15{\sim}40$ m/s(Pitot Tube)、 $0.15{\sim}15$ m/s(Velocity Grid)
Air velocity	Accuracy	±3% of readings ±0.05m/s(>0.25m/s)
	Resolution	0.01m/s
Differential	Test range	-2500~2500 Pa
	Accuracy	±0.25% of readings ±1Pa
pressure	Resolution	0.001 Pa
	Test range	0~60°C
Temperature	Accuracy	±0.5℃
·	Resolution	0.1℃
	Test range 0~100%RH	
Humidity	Accuracy	±3%RH(10~90%RH)
	Resolution	0.1%RH
Absolute	Test range 70~130 kPa	
	Accuracy	±2% of readings
pressure	Resolution	0.1 kPa
Operation Temp.		0~60°C (non condensing )
Storage Temp.		-20~70°C (non condensing)
Power Source		4 pcs of AA-size buttery (approx. 9hrs available) or DC5V Adapter
Weight		3.6kg approx.




### 2.1. Base structure


#### 1. Outlook




# 2. Internal view



3. Grid






# 3. Settubg up

# 3.1. Capture Hood

1. Fabric Hood & basement




#### Note:

- 1) Make sure the elastic band of the fabric hood is placed exactly as shown (Figure (3)) to ensure a good fit. The elastic band must lineup with the top lip of the base as shown. If you don't line it up correctly, the hood might slide off the base when you insert the poles.
- 2) Use the 4 positioning holes on the base stitching as a guide to ensure the hood fully wraps around the base..
- 2. Frame Poles




Note: The poles always cross in an "X" shape when assembling. To remove the poles, simply reverse the steps.

#### 3. Potable handle



Note: Release the poles and portable handle when packing in the carring case to avoid case damaged.

#### 4. Micromanometer & basement




Note: 1) Correct install the micromanometer to the base with the same S/N No.

2) Do not cross-insert the Air pines just refer to the figure as above.

# 3.2. Velocity Grid


Air velocity testing can be achieved when micromanometer work together with the velocity Grid.



Note: Mounting angle between the Grid and handle can be adjusted as request.


### 3.3. Pitot tube

Air flow and velocity testing can be achieved when micromanometer work together with Pitot tube.




# 3.4. Micromanometer

Differential pressure testing can be achieved by using the micromanometer.



# 3.5. Shoulder Strap


When testing by Velocity Grid、Pitot tube、Micromanometer as the tool, for convenient view and easy operation, the shoulder strap can be fit for using.




Ti Operating motivation

# 4.1. Power supply


1. Power by AC adapter. AC adapter will be as the priority power supply when the AC adapter and batteries are all available using. The specification of the AC adapter is: I/P:AC 110-240V 50/60HZ O/P:DC 5V/2A.



- 2. Power by battery
- ◆ 4 AA-size batteries can be used for a power supply.
- ◆ Press the compartment cover lock (refer to figure indicating ①/②), and then slide the cover off to open the battery compartment。
- Put 4 AA-size batteries in and make certain the batteries are correctly oriented.
- ◆ New alkaline battery or rechargeable Ni-MH battery can be used. Never mix battery types, or battery leakage or damage to the instrument may occur.
- Replace the compartment cover (refer to figure ③). The compartment cover will be automatically locked.




◆ When power is supplied with batteries, the current charge value will be displayed on the upper of the LCD. When the batteries run out of power, the instrument will turn off.



#### Note:

- 1) Do not test with low battery power.
- 2) Do not mixing use full power battery with low power ones for ensure testing well.
- 3) Change the batteries in time when power shows low for ensure testing well.

# 4.2. Start On / Power Off







1. Start On

Press button [POWER] for 2 seconds for start On, displaying "KANOMAX" then enter into the testing interface.


2. Power Off


Press [POWER] for 2 seconds, shut down and power off.

# 4.3. Keypad operation

Keypad panel includes control buttons, back-pressure switch and LCD adjusting keys.

1. Control buttons:





UP: Value increasing or Cursors upward moving.

SAVE: Save the current test results

SET: Store changes or enter into the selected option's next level interface

MENU: On the main screen holding this down for 2 seconds will enter into the main "MENU"

OK: In the "General Setting", pressing for 2 seconds will save the date or time;

While in the "Record Processing", pressing for 2 seconds will delete saved data.

START: Begin a measurement in main menu

STOP: Stop a measurement in main menu

BACK: Cancel or end an operation or return to the previous screen

EXIT: Return to the main menu


ESC: Stop printing when printing

DOWN:: Value decreasing or Cursors downward moving


2. Back pressure switch for on/off

#### Note:

- 1) If you are not in the back pressure mode, shut down the back pressure plate and lock the switch to avoid testing errors.
- 2) If the airflow is greater than 1500m³/h (883CFM), testing in back-pressure Mode is recommended according to testing procedures.



#### 3. LCD adjusting



#### Note:


- 1) Please switching the handle slow gently during LCD adjusting.
- 2) Stop turning when reach to the bound, for avoiding over turning and handle falling down.

### 5. Function test

Besides the airflow test, if along with other optional products, test air velocity and micro-differential pressure is available.

#### 5.1. User Interface

Testing information, parameter settings and the testing conditions will be displayed on the main interface.



- A. Date and time —— real-time display the current date and time
- B. Testing data display —— airflow , air velocity , differential pressure and etc.
- C. Tools icon working tools display
- D. Testing condition —— temperature , humidity or absolute pressure displaying according to the

- E. Keypad —— current status indicating and the grey state is not available.
- F. Testing mode —— display the current testing mode
- G. Cycle—recorded Cycle quantity
- H. ID—— current ID quantity
- I. Power supply —— AC adapter or batteries
- J. Blue tooth status —— when blue tooth working, the icon display or no icon display with closed.


### 5.2. Single Mode



#### Note:

- 1) START: Left-button for the Micromanometer or instrument base
- 2) After testing start for about 8s with stably measuring, output displaying the test data and testing wind direction.
- 3) Testing stop when finish outputting.
- 4) Press SAVE for data record with adding 1 to the Cycle value.
- 5) Same testing method with the capture hood when using the Velocity Grid、Pitot、Micromanometer.

## 5.3. Average Mode




#### Note:

- 1) START: Left-button for the Micromanometer or instrument base.
- 2) After testing start for about 8s with stably measuring, output displaying the test data for the first time.
- 3) Go on the testing with frequency of 1s for data updating.
- 4) The continuously outputting is as the average data and be related with the "AvgTime" set by user. More larger of the value of "AvgTime" will get more stable of the testing data.
- 5) If auto-save function On, testing data will be recorded automatically according to the setting "AutoSave" and value of "Cycle" will be added automatically.
- 6) Press "STOP", testing stop. And "STOP" will be as the Left-button for the Micromanometer or instrument base.
- 7) Same testing method with the capture hood when using the Velocity Grid、Pitot、Micromanometer.

#### 5.4. Back Pressure Mode

get the more real air flow, we suggest testing under the Back Pressure Mode when air flow more than 1500 m<sup>3</sup>/h.



#### Note:

- 1) START: Left-button for the Micromanometer or instrument base.
- 2) After testing start for about 8s, indicating "Open Flap". It's time for switching on the Back-Pressure Plate as needed.
- 3) When Back-Pressure Plate opened On, tip of "Open Flap" disappear.
- 4) Press START for about 8s, testing stop, real airflow outputting with back pressure compensation.
- 5) Press SAVE for testing data record and the meanwhile, value Cycle added 1.
- 6) When select Velocity Grid、Pitot tube、Micromanometer for testing, Back-Pressure compensation will be not available.


# 6. Menu setting

when stop for ready, press for 2s for enter into the menu.






#### 6.1.1. Date




#### 6.1.2. Time




#### 6.1.3. Auto Off

When instrument not in use, it will be turn off automatically according to Auto-Off time setting to avoid power waste.



### 6.1.4. Backlighting

Adjust the LCD lighting according to the environmental brightness for the best viewing. And meanwhile, the lower the brightness, the less battery power will be needed.




#### 6.1.5. Communication


Communicate with PC, printer and Bluetooth are available. Data transfer or remote monitoring is available by selecting communicating way.

Note:


- PC communication means to save the testing record to the computer. And the communication cable and software reserved for data-fetching are necessary for using together. Refer to the << Manuel for Data fetching Software>> .
- 2) For printing out the testing record, the optional printer and cables for special are necessary. Printer baud rate is set to 19200 bps and Do not change it. Refer to the << Manuel for RD series Mini printer >> .
- 3) Bluetooth make the remote control and data monitoring available by Hand phone. And our APP software is necessary for using together. Refer to the << Manuel for TAB Mobile>>.
- 4) Shut off the switch when not use the Bluetooth communication for improving battery life.



5) If other settings need to be changed when Bluetooth successful communication, you need to shutdown Bluetooth firstly then enter the menu to change other settings. Just as below shown:




# 6.2. Test setting




6.2.1. Tools


measurements. The optional accessories (Velocity Grid, Pitot tube, and Micromanometer) expand the parameters that the instrument can measure. Refer to the Chapter 3 and 5 for details.




#### 6.2.2. ID application




#### 6.2.3. Test Mode





#### 6.2.5. Data record




#### Note:

- Browsing or deleting test data in current testing tools is available for allowed.
- 2) Data exporting by communication function and deleting the non-valid data in time for the effective storage space.
- 3) Pressing for 2 seconds will delete saved data.


### 6.2.6. Standard/Actual test



- 1) Standard Test (Std): in the condition of atmosphere 101.3kPa and temperature 21.1°C.
- 2) Actual Test(Act): in actual condition with the compensation for the atmosphere and temperature.
- 3) No Std/Act selection when in Micromanometer testing.
- 4) When doing actual testing under Velocity Matrix and pitot tube, manually input the current temperature is needed for the compensation. Refer to the figures as below:



#### 6.2.7. Compensation factor



#### Note:

1) Under Capture Hood and Velocity Grid testing, K factor is used for testing data compensation in order to get more accuracy result. And the testing range will be  $0.500 \sim 1.500$ .

Note: K factor will affect the accuracy of measurement, so, it's necessary for having the test bench for standard reference and using with caution.

#### For example:

If the Capture Hood is as the selected testing tool, testing value by Model 6715 is  $1000 \text{m}^3/\text{h}$ , the standard airflow is  $1020 \text{m}^3/\text{h}$ , at this point, K factor= standard value/testing value =1.020, input this K factor to correct the accuracy of the measurement by Model 6715.

- 2) Under Pitot tube testing, K factor is the factory standard coefficient and each pitot with its own independent factor.
- 3) No K factor setting under the micromanometer testing.

Symptom	Possible causes	Corrective action
	incorrect specification of the AC	Check and refer to Chapter4.1
	Adapter	in Manuel
No display when power On	Datter installation ower	Check and refer to Chapter4.1
	Battery installation error	in Manuel
	Low or dead battery power	Replace the batteries
Display "Connect Error"	Connection cable loss with the	Reseat the connect cable
Display Connect Life	base	Neseat the connect cable
Display "DATA Error"	Back-Pressure mode using	Refer to Chapter 5.4 in Manuel
Display DATA ETTO	incorrect	Kerer to Chapter 3.4 in Manuer
Press "START" but no response	Low battery charge	Replace new batteries
Tress START but no response	Back Pressure Plate opened On	Shut off the plate
Testing record can not be saved	Over-limit Cycle quantity in	Add new ID or delete some
resting record carriot be saved	current ID	data
Navy ID as a wat has a dated	ID over-limit	Delete some ID
New ID can not be added	Current ID is without "Cycle"	Using the current ID
	Unmatched printer	Using our special optional
Printing unreadable code	Offinatorieu printer	printer supplied by us.
Trinking unleadable code	Wrong baud rate setting	Reset referring to the Printer
	With one badd rate setting	Manuel

# 8 Warranty and Service

# 8.1. Product Warranty

The limited warranty set forth below is given by KANOMAX GROUP COMPANIES with respect to the KANOMAX brand capture hood and other accessories (hereafter referred to as "PRODUCT") purchased directly from KANOMAX GROUP COMPANIES or from an authorized distributor.

Your PRODUCT, when delivered to you in new condition in its original container, is warranted against defects in materials or workmanship as follows: for a period of one (1) year from the date of original purchase, defective parts or a defective PRODUCT returned to KANOMAX GROUP COMPANIES, as applicable, and proven to be defective upon inspection, will be exchanged for a new or comparable rebuilt parts, or a refurbished PRODUCT as determined by KANOMAX GROUP COMPANIES. Warranty for such replacements shall not extend the original warranty period of the defective PRODUCT.

This limited warranty covers all defects encountered in normal use of the PRODUCT, and does not apply in the following cases:

- (1) Use of parts or supplies other than the PRODUCT sold by KANOMAX GROUP COMPANIES, which cause damage to the PRODUCT or cause abnormally frequent service calls or service problems.
- (2) If any PRODUCT has its serial number or date altered or removed.
- (3) Loss of damage to the PRODUCT due to abuse, mishandling, alternation, improper packaging by the owner, accident, natural disaster, electrical current fluctuations, failure to follow operation, maintenance or environmental instructions prescribed in the PRODUCT's operation manual provided by KANOMAX GROUP COMPANIES, or service performed by other than KANOMAX GROUP COMPANIES. NO IMPLIED WARRANTY, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE, APPLIES TO THE PRODUCT AFTER THE APPLICABLE PERIOD OF THE EXPRESS LIMITED WARRANTY STATED ABOVE, AND NO OTHER EXPRESS WARRANTY OR GUARANTY, EXCEPT AS MENTIONED ABOVE, GIVEN BY ANY PERSON OR ENTITY WITH RESPECT TO THE PRODUCT SHALL BIND KANOMAX GROUP COMPANIES. KANOMAX GROUP COMPANIES SHALL NOT BE LIABLE FOR LOSS OF STORAGE CHARGES, LOSS OR CORRUPTION OF DATA, OR ANY OTHER SPECIAL, INCIDENTAL OR CONSEQUENTIAL DAMAGES CAUSED BY THE USE OR MISUSE OF, OR INABILITY TO USE, THE PRODUCT, REGARDLESS OF THE LEGAL THEORY ON WHICH THE CLAIM IS BASED, AND EVEN IF KANOMAX GROUP COMPANIES HAS BEEN ADVISED OF THE POSSIBILITY OF SUCH DAMAGES. IN NO EVENT SHALL RECOVERY OF ANY KIND AGAINST KANOMAX GROUP COMPANIES BE GREATER IN AMOUNT THAN THE PURCHASE PRICE OF THE PRODUCT SOLD BY KANOMAX GROUP COMPANIES AND CAUSING THE ALLEGED DAMAGE. WITHOUT LIMITING THE FOREGOING, THE OWNER ASSUMES ALL RISK AND LIABILITY FOR LOSS, DAMAGE OF, OR INJURY TO THE OWNER AND THE OWNER'S PROPERTY AND TO OTHERS AND THEIR PROPERTY ARISING OUT OF USE OR MISUSE OF, OR INABILITY TO USE, THE PRODUCT NOT CAUSED DIRECTLY BY THE NEGLIGENCE OF KANOMAX GROUP COMPANIES. THIS LIMITED WARRANTY SHALL NOT EXTEND TO ANYONE OTHER THAN THE ORIGINAL PURCHASER OF THE PRODUCT, OR THE PERSON FOR WHOM IT WAS PURCHASED AS A GIFT, AND STATES THE PURCHASER'S EXCLUSIVE REMEDY.

#### 8.2. After service

- ◆ When you have a problem with your instrument, please check out the "Common Trouble Shooting" section first
- ◆ If that does not help, please contact your local distributor, or contacts on the last page.
- ◆ During the warranty period, we will repair at no charge a product that proves to be defective due to material or workmanship under normal use. (See Section 14.1 Warranty)
  - All return shipping charges are the responsibility of the customer.

- Upon request, we will repair the instrument at the customer's expense, if the instrument's performance is found to be recoverable by providing the repair.
- Replacement parts are available for a minimum period of five (5) years after termination of production. This storage period of replacement parts is considered as the period during which we can provide repair service. For further information, please contact your local distributor, or contacts on the last page.

When making an inquiry, please provide the following information:			
* PRODUCT NAME			
* Model No.			
* Serial No.			
* Description of the problem :			
* Data of Purchase: Day, Month and Year			