



BROMINE KIT

OCTA-SLIDE 2, 0.2-3.0 ppm


CODE 6955-01

QUANTITY	CONTENTS	CODE
50	Bromine Tablets	6977A-H
2	Test Tubes, plastic, 2.5-5-10mL, w/caps	0106
1	Octa-Slide 2 Viewer	1101
1	Bromine Octa-Slide 2 Bar, 0.2-3.0 ppm	6822-01

*WARNING: Reagents marked with an * are considered to be potential health hazards. To view or print a Material Safety Data Sheet (MSDS) for these reagents go to www.lamotte.com. To obtain a printed copy, contact LaMotte by e-mail, phone or fax.

To order individual reagents or test kit components, use the specified code number.







USE OF THE OCTA-SLIDE 2 VIEWER



The Octa-Slide 2 Viewer should be held so non-direct light enters through the back of the Viewer. Slide the Octa-Slide 2 Bar into the Viewer. Insert the reacted sample into the top of the Viewer. Match the color of the reaction to the color standards.

WARNING! This set contains chemicals that may be harmful if misused. Read cautions on individual containers carefully. Not to be used by children except under adult supervision.

PROCEDURES

<p>1</p>  <p>Insert Bromine Octa-Slide 2 Bar (6822-01) into the Octa-Slide 2 Viewer (1101).</p>	<p>2</p>  <p>Fill a test tube (0106) to the 5 mL line with the water sample.</p>	<p>3</p>  <p>Add one Bromine Tablet (6977A).</p>
<p>4</p>  <p>Cap and mix until tablet disintegrates.</p>	<p>5</p>  <p>Insert test tube into Octa-Slide 2 Viewer.</p>	<p>6</p>  <p>Match sample color to a color standard. Record as ppm Bromine.</p>

NOTE: Thoroughly clean and rinse test tubes after each use.