
- Single Hole Mounting
- 2-Wire Operation
- 4 to 25V d.c. Full Scale Reading
- 12.5mm / 0.5" LCD Digit Height
- Reverse Polarity Protection
- Splashproof






Easy Mounting Voltmeter	Stock Number EMV 1200			
Specification	Min.	Тур.	Max.	Unit
Accuracy (overall error)		0.2		V (±1 count)
Linearity			±1	count
Valid displayed reading*	4.0		25.0	V d.c.
Resolution		100		mV d.c.
Sample rate		3		sample/sec
Operating temperature range	0		50	°C
Supply voltage	3		50	V d.c.
Supply current			3	mA

- * At voltages below 4.0V d.c., the LCD will display LO and will ultimately go blank. At Voltages in excess of 25.0V d.c., the LCD will display HI.
- Do NOT exceed 50V d.c., as this may damage the meter.
- The module includes reverse polarity protection up to 30V for up to 30 seconds.

Operation and accuracy at voltages above 25.0V or below 4.0V are not specified.

SAFETY

To comply with the Low Voltage Directive (LVD 93/68/EEC), input voltages to the module's wires must not exceed 60Vd.c. The user must ensure that the incorporation of the meter into the user's equipment conforms to the relevant sections of BS EN 61010 (Safety Requirements for Electrical Equipment for Measuring, Control and Laboratory Use).






LASCAR ELECTRONICS LTD. MODULE HOUSE, WHITEPARISH,WILTSHIRE, SPS 251. LIK

LASCAR ELECTRONICS INC. 4258 WEST 12th STREET, ERIE, PA 16505, USA LASCAR ELECTRONICS (HK) LIMITED
UNIT NOS. 6-8, 19/F FUTURA PLAZA,
111-113 HOW MING STREET, KWUN TONG,
KOWLOON, HONG KONG

1.888.610.7664

www.calcert.com

AUTOMOTIVE AND MARINE PROTECTION



Automotive and maritime applications can create harsh conditions for electronic devices. Fig. 1 shows the type of protection we recommend is added when the EMV 1200 is used in an application of this nature. We suggest a high wattage 33 ohm resistor is used in series with two 5W zener diodes where the value of each zener is at least 3V higher than the maximum you intend to measure. Alternatively it is possible to use a single 1500W transient voltage suppressor (TVS) in place of the zener diodes. In addition it is recommended that the meter is connected as close to the battery as is convenient.

Specifications liable to change without prior warning

EMV 1200

Issue 10 October/2008

IRCN 695

S.C.

Applies to EMV 1200/2

